§ 888.3230

weight polyethylene, and is limited to those prostheses intended for use with bone cement (§ 888.3027).

- (b) Classification. Class III.
- (c) Date PMA or notice of completion of a PDP is required. A PMA or a notice of completion of a PDP is required to be filed with the Food and Drug Administration on or before December 26, 1996 for any finger joint metal/polymer constrained cemented prosthesis that was in commercial distribution before May 28, 1976, or that has, on or before December 26, 1996 been found to be substantially equivalent to a finger joint metal/polymer constrained cemented prosthesis that was in commercial distribution before May 28, 1976. Any other finger joint metal/polymer constrained cemented prosthesis shall have an approved PMA or a declared completed PDP in effect before being placed in commercial distribution.

[52 FR 33702, Sept. 4, 1987, as amended at 61 FR 50709, Sept. 27, 1996]

§888.3230 Finger joint polymer constrained prosthesis.

- (a) *Identification*. A finger joint polymer constrained prosthesis is a device intended to be implanted to replace a metacarpophalangeal or proximal interphalangeal (finger) joint. This generic type of device includes prostheses that consist of a single flexible acrossthe-joint component made from either a silicone elastomer or a combination pf polypropylene and polyester material. The flexible across-the-joint component may be covered with a silicone rubber sleeve.
 - (b) Classification. Class II.

§888.3300 Hip joint metal constrained cemented or uncemented prosthesis.

(a) Identification. A hip joint metal constrained cemented or uncemented prosthesis is a device intended to be implanted to replace a hip joint. The device prevents dislocation in more than one anatomic plane and has components that are linked together. This generic type of device includes prostheses that have components made of alloys, such as cobalt-chromium-molybdenum, and is intended for use with or without bone cement (§888.3027).

This device is not intended for biological fixation.

- (b) Classification. Class III.
- (c) Date PMA or notice of completion of a PDP is required. A PMA or a notice of completion of a PDP is required to be filed with the Food and Drug Administration on or before December 26, 1996 for any hip joint metal constrained cemented or uncemented prosthesis that was in commercial distribution before May 28, 1976, or that has, on or before December 26, 1996 been found to be substantially equivalent to a hip joint cemented constrained metal uncemented prosthesis that was in commercial distribution before May 28, 1976. Any other hip joint metal constrained cemented or uncemented prosthesis shall have an approved PMA or a declared completed PDP in effect before being placed in commercial distribution.

 $[52\ FR\ 33702,\ Sept.\ 4,\ 1987,\ as\ amended\ at\ 61\ FR\ 50709,\ Sept.\ 27,\ 1996]$

§ 888.3310 Hip joint metal/polymer constrained cemented or uncemented prosthesis.

- (a) Identification. A hip joint metal/ polymer constrained cemented uncemented prosthesis is a device intended to be implanted to replace a hip joint. The device prevents dislocation in more than one anatomic plane and has components that are linked together. This generic type of device includes prostheses that have a femoral component made of alloys, such as cobalt-chromium-molybdenum, and an acetabular component made of ultrahigh molecular weight polyethylene. This generic type of device is intended for use with or without bone cement (§888.3027). This device is not intended for biological fixation.
 - (b) Classification. Class III.
- (c) Date PMA or notice of completion of a PDP is required. A PMA or a notice of completion of a PDP is required to be filed with the Food and Drug Administration on or before December 26, 1996 for any hip joint metal/polymer constrained cemented or uncemented prosthesis that was in commercial distribution before May 28, 1976, or that has, on or before December 26, 1996 been found to be substantially equivalent to a hip

joint metal/polymer constrained cemented or uncemented prosthesis that was in commercial distribution before May 28, 1976. Any other hip joint metal/polymer constrained cemented or uncemented prosthesis shall have an approved PMA or a declared completed PDP in effect before being placed in commercial distribution.

 $[52\ FR\ 33702,\ Sept.\ 4,\ 1987,\ as\ amended\ at\ 61\ FR\ 50710,\ Sept.\ 27,\ 1996]$

§888.3320 Hip joint metal/metal semiconstrained, with a cemented acetabular component, prosthesis.

- (a) *Identification*. A hip joint metal/ metal semi-constrained, with a cemented acetabular component, prosthesis is a two-part device intended to be implanted to replace a hip joint. The device limits translation and rotation in one or more planes via the geometry of its articulating surfaces. It has no linkage across-the-joint. This generic type of device includes prostheses that consist of a femoral and an acetabular component, both made of alloys, such as cobalt-chromium-molybdenum. This generic type of device is limited to those prostheses intended for use with bone cement (§888.3027).
 - (b) Classification. Class III.
- (c) Date PMA or notice of completion of a PDP is required. No effective date has been established of the requirement for premarket approval. See §888.3.

§888.3330 Hip joint metal/metal semiconstrained, with an uncemented acetabular component, prosthesis.

- (a) Identification. A hip joint metal/ metal semi-constrained, with uncemented acetabular component, prosthesis is a two-part device intended to be implanted to replace a hip joint. The device limits translation and rotation in one or more planes via the geometry of its articulating surfaces. It has no linkage across-the-joint. This generic type of device includes prostheses that consist of a femoral and an acetabular component, both made of alloys, such as cobalt-chromium-molybdenum. The femoral component is intended to be fixed with bone cement. The acetabular component is intended for use without bone cement (§888.3027).
 - (b) Classification. Class III.

(c) Date PMA or notice of completion of a PDP is required. No effective date has been established of the requirement for premarket approval. See §888.3.

§ 888.3340 Hip joint metal/composite semi-constrained cemented prosthesis.

- (a) Identification. A hip joint metal/ composite semi-constrained cemented prosthesis is a two-part device intended to be implanted to replace a hip joint. The device limits translation and rotation in one or more planes via the geometry of its articulating surfaces. It has no linkage across-the-joint. This generic type of device includes prostheses that consist of a femoral component made of alloys, such as cobaltchromium-molybdenum, and an acetabular component made of ultra-high molecular weight polyethylene with carbon fibers composite. Both components are intended for use with bone cement (§888.3027).
 - (b) Classification. Class II.

§ 888.3350 Hip joint metal/polymer semi-constrained cemented prosthesis.

- (a) Identification. A hip joint metal/ polymer semi-constrained cemented prosthesis is a device intended to be implanted to replace a hip joint. The device limits translation and rotation in one or more planes via the geometry of its articulating surfaces. It has no linkage across-the-joint. This generic type of device includes prostheses that have a femoral component made of alloys, such as cobalt-chromium-molybdenum, and an acetabular resurfacing component made of ultra-high molecular weight polyethylene and is limited to those prostheses intended for use with bone cement (§888.3027).
 - (b) Classification. Class II.

§ 888.3353 Hip joint metal/ceramic/ polymer semi-constrained cemented or nonporous uncemented prosthesis.

(a) *Identification*. A hip joint metal/ceramic/polymer semi-constrained cemented or nonporous uncemented prosthesis is a device intended to be implanted to replace a hip joint. This device limits translation and rotation in one or more planes via the geometry of its articulating surfaces. It has no